

Reducing scan time using time-of-flight protocol in clinical [¹⁸F]FDG-PET/CT imaging: A feasibility study

Roya Sharifpour^{1,2}, Pardis Ghafarian^{3,4}, Mehrdad Bakhshayesh-Karam^{3,4},
Carlos F. Uribe^{5,6}, Arman Rahmim^{6,7}, Mohammad Reza Ay^{1,2}

¹Department of Medical Physics and Biomedical Engineering, Tehran University of Medical Sciences, Tehran, Iran

²Research Center for Molecular and Cellular Imaging, Tehran University of Medical Sciences, Tehran, Iran

³Chronic Respiratory Diseases Research Center, National Research Institute of Tuberculosis and Lung Diseases, Shahid Beheshti University of Medical Sciences, Tehran, Iran

⁴PET/CT and Cyclotron Center, Masih Daneshvari Hospital, Shahid Beheshti University of Medical Sciences, Tehran, Iran

⁵Department of Functional Imaging, BC Cancer Research Institute, Vancouver, Canada

⁶Departments of Radiology and Physics, University of British Columbia, Vancouver, Canada

⁷Department of Integrative Oncology, BC Cancer Research Institute, Vancouver, Canada

Corresponding Author:

Pardis Ghafarian, PhD

Chronic Respiratory Diseases Research Center, National Research Institute of Tuberculosis and Lung Diseases (NRITLD), Shahid Beheshti University of Medical Sciences, Tehran, Iran

Email: pardis.ghafarian@sbm.ac.ir

Short title: Reducing [¹⁸F]FDG PET/CT scan time using TOF protocol

ABSTRACT

Introduction: This study evaluates scan time reduction using time-of-flight (TOF) PET, when quantitative parameters including volumetric measures are considered.

Methods: 32 patients were included in the study. The reconstruction parameters for TOF were 2 iterations, 18 and 24 subsets, and for non-TOF was 3 iterations and 18 subsets. A post smoothing filter with FWHM of 5.4 mm and 6.4 mm were used for TOF and 6.4 mm for non-TOF. TOF reconstruction was performed with 2, 2.5 and 3 min/bed position, and 3 min/bed position scan time was applied in non-TOF. Quantitative parameters such as coefficient of variation (COV), signal-to-noise ratio (SNR), lesion-to-background ratio (LBR), metabolic tumor volume (MTV) and total lesion glycolysis (TLG) were utilized. Standard uptake value (SUV) was also measured. Different segmentation thresholds were studied.

Results: Improvement in SNR for TOF relative to non-TOF was observed when utilizing 18 subsets, 5.4 mm filter size with 3 min scan time/bed position (P-value<0.0001), and for 18 subsets, 6.4 mm filter size when 2.5- and 3-min scan time/bed positions was applied (P-value≤0.02). Scan time reduction did not illustrate significant variation for the SUVs and lesion size. In all TOF protocols for both TLG and MTV, the measured values decreased with increasing segmentation thresholds, as expected, with significantly more impact for higher thresholds (70%, 75%). Meanwhile, higher values were observed for higher post smoothing filter in each specified threshold. With increasing of the threshold, ΔTLG was increased with

more impact for higher post smoothing filter. Δ MTV were -10.10 ± 11.09 (-1.35 ± 8.59) and 0.68 ± 17.51 (12.42 ± 18.90) in 2 min/bed position with 5.4 (6.4 mm post smoothing filter) for threshold of 45% and 75% respectively.

Conclusion: Scan time reduction from 3 to 2 min can be obtained with TOF in comparison with non-TOF, especially when higher segmentation threshold values with higher subset number (24) and 6.4 mm filters are utilized.

Key Words: [^{18}F]FDG PET/CT; Scan time; Time-of-flight; Metabolic tumor volume; Total lesion glycolysis

INTRODUCTION

Positron emission tomography/computed tomography (PET/CT) with ^{18}F -2-fluoro-2-deoxy-D-glucose ([^{18}F]FDG) has been used in diagnosis and staging evaluation of treatment response, disease recurrence and radiotherapy treatment planning. Today, commercially available PET/CT scanners enable modeling of the point spread function (PSF), aiming to improve the spatial resolution of images [1, 2] and allowing reduction of the partial volume effect (PVE) in PET images [3-5]. Using time-of-flight (TOF) information in the reconstruction can improve the localization of annihilation photons [6, 7] to increase signal-to-noise ratio (SNR) [8] and contrast [9, 10]. These effects are with more impact in clinical images of heavy patients and low count rate data [11-13].

Previous study showed that TOF can also decrease respiratory artifacts [9]. Reduction of scan time and/or injection of lower activity can be also seen for PET images reconstructed with TOF method [13, 14]. SUVmean, SUVmax and SUVpeak are common semi-quantitative parameters used for quantification of PET images [15]. It was seen that low count rates can lead to upward bias in SUVmean up to 15% due to decrease in SNR [16]. The bias is more severe for SUVmax owing to increase in noise [5, 17]. Today, SUVmax is a popular index for evaluation of [^{18}F]FDG uptake in PET/CT imaging. As SUVs indicate to one or some voxels in defined tissue, it may not introduce correct findings of metabolic activity [18]. In recent years, significant improvement and development is performed for quantification of PET/CT images in the application of treatment planning and treatment response evaluation [19]. It is observed that volumetric parameters obtained with PET/CT imaging such as metabolic tumor volume (MTV) and total lesion glycolysis (TLG) can deliver important information for tumor burden and prognosis findings in evaluation of patient treatment response [20-22]

Fixed thresholding based on SUVmax has been applied in biological target delineation and patient treatment response [23, 24]. It is illustrated that various reconstruction methods and PET/CT segmentation techniques in clinical PET/CT can also produce an important discrepancy in MTV and tumor delineation [25, 26]. One of the main challenges in PET-based quantitative purpose especially for patient treatment response depends on the technical issues such as scan time, reconstruction methods and reconstruction parameters, which affect the uptake value, target size and target shape [27, 28]. Albano et al. [29] stated that percentage of variation for MTV and TLG with consideration of baseline and end of treatment are strongly correlated to progression-free survival and overall survival for patient with Burkitt's lymphoma. Ketabi et al. revealed that the relative difference of SUVs (SUVmax and SUVpeak) in high and low background activity was more important for the lesion sizes with less than four times the FWHM of the scanner resolution versus other lesion sizes [18]. However, this behavior was seen with less impact for SUVpeak. It should be noted that image quality can be also changed by lesion to background ratio, lesion activity and background activity separately [30]. In the study performed by Sharma et al., SUVmax, SUVmean, and tumor-to-background ratio of baseline PET/CT scan did not illustrate strong behavior as prognostic indicators for overall survival of patients with non-small cell lung cancer that performing platinum-based

chemotherapy, that was not in agreement with their finding for MTV and TLG [31]. Gencturk et al. revealed that PET/CT imaging prior to treatment for patients with salivary gland adenoid cystic carcinoma with consideration of metabolic parameters may be useful for predicting distant metastasis-free survival, progression-free survival and overall survival [32]. Ye et al. also have shown that metabolic volumetric parameters (MTV and TLG) correlate with the level of preoperative serum cancer antigen 125 (CA 125), so might be predictors for survival in ovarian CCC patients [33].

Optimum reconstruction parameters are also critical in quantitative analysis [5, 34, 35]. Reduction of scan time has the potentials to increase patient throughput and to reduce motion induced artifacts in images. In previous investigations the range of optimum reconstruction parameters was selected according to our patients' data [9, 18, 19, 36].

The aim of this study was to evaluate the impact that reducing scan time has on the quantitative metabolic and volumetric parameters. The contribution of scan time reduction was assessed when combination of TOF and PSF protocols were considered for clinical data. The patient's administered activity was not changed (4.6 MBq/kg).

METHODS

Patients

Thirty-two patients (23–78 years) including 19 men and 13 women with lesions located in thorax area with range of 18.31–42.70 mm in FWHM; 1.60–6.76 in signal to background ratio (SBR) (defined as SUV_{max} in lesion to SUV_{mean} in normal liver); as measured in our routine protocol, were retrospectively assessed. The patient data were classified as follow: 5 lung cancer, 4 non-Hodgkin lymphoma, 3 kidney cancer, 3 Hodgkin lymphoma, 3 colon cancer, 4 renal cell carcinoma, 1 pancreatic cancer, 1 gastric cancer, 3 breast cancer, 4 esophageal cancer, and 1 unknown primary. Patients with fasting blood sugar level higher than 200 mg/dL were not included in this study. After 6–8 hours fasting period, 4.6 MBq/kg of [¹⁸F]FDG was injected intravenously according to European Association of Nuclear Medicine (EANM) guidelines [37]. Patient's BMI were 24.04±2.01 with weight range of 58–89 kg. Patients were scanned after 61.1±1.3 minutes uptake time and asked to void before imaging procedure.

Data acquisition and image reconstruction

All data was acquired on Discovery 690 VCT (GE Healthcare, Milwaukee, USA) equipped with 64-slice CT (Light Speed VCT). The PET system consists of 24 detector rings, covering 15.7 cm and 70 cm axial and trans-axial field-of-view (FOV), respectively. The PET scanner comprises a total of 13,824 lutetium-yttrium oxyorthosilicate (LYSO) crystals with dimension of 4.2×6.3×25 mm³. The coincidence time window is 4.9 ns and timing resolution is 555 ps.

First, the CT scan was performed for attenuation correction and localization purposes from head to mid-thigh, using the automatic current modulation method. Next, the emission data were collected in list mode with 3 min/bed position. Both scatter and attenuation correction were applied on FDG-PET raw data which were reconstructed with two algorithms of TOF (OSEM +PSF+TOF) and non-TOF (OSEM +PSF; routine protocol in our department with 3 min per bed position). The reconstruction parameters for TOF protocol were 2 iterations, 18 and 24 subsets, and those for routine protocol (non-TOF) were 3 iterations and 18 subsets. A Gaussian filter as a post smoothing filter with a full width at half maximum (FWHM) of 5.4 mm and 6.4 mm were used for TOF and those for non-TOF was 6.4 mm. All parameters used in TOF protocols were optimum, based on our previous studies [9, 18, 19, 36]. All data were reconstructed into a 256×256 image matrix. In order to evaluate the effect of scan time reduction on PET images, the list mode data were reconstructed with 2, 2.5 and 3 min per bed position in all TOF protocols.

Assessment strategy

The impact of scan time reduction in all TOF protocols versus non-TOF protocol was evaluated by various quantitative and semi quantitative analysis in PET/CT images. Coefficient of variation (COV), signal-to-noise ratio (SNR), and lesion-to-background ratio (LBR) were calculated. 9 ROIs with 30 mm in diameter were depicted on the 3 axial slices in the largest and uniform liver section and around this section (3 ROIs in each slice) with regard not to include the porta-hepatis and major vessels; COV was evaluated according to the equation 1:

$$COV(\%) = \frac{SD_{Liver}}{C_{Liver}} \times 100 \quad (1)$$

where C_{Liver} is the average activity concentration and SD_{Liver} is the average of standard deviations in 9 ROIs.

SNR was also measured according to the equation 2:

$$SNR = \frac{C_{Lesion} - C_{Liver}}{SD_{Liver}} \quad (2)$$

where C_{Lesion} is the max activity concentration in lesion.

In addition, LBR was assessed according to the equation 3:

$$LBR = \frac{C_{Lesion}}{C_{Liver}} \quad (3)$$

Semi-quantitative analysis consisted of SUVmax, SUV50% (using 50% thresholding with respect to SUVmax) was performed. In addition, total lesion glycolysis (TLG) with various thresholds (45%, 50%, 70% and 75%) was also measured according to the equation 4:

$$TLG = SUV_{mean} \times MTV \quad (4)$$

where SUVmean is the average of SUV in thresholding value that corresponding to metabolic tumor volume (MTV).

The effect of scan time reduction on lesion size was also assessed by calculation of maximum lesion diameter. In order to calculate the maximum diameter (called lesion size), the lesion profile was drawn in the direction in which the lesion had the largest size (the direction was chosen on the non-TOF image). Then, an appropriate Gaussian curve was fitted. The full width at half maximum of Gaussian curve was considered as the estimation of maximum lesion diameter [9].

The relative differences for TLG, MTV, SUV and lesion size between TOF and non-TOF protocols were calculated according to the equation 5:

$$\Delta MTV / \Delta TLG / \Delta SUV / \Delta lesion_size = \frac{Value_{TOF} - Value_{non-TOF}}{Value_{non-TOF}} \times 100 \quad (5)$$

It should be noted that visual qualitative assessment of image quality for different scan time was also performed by two expert physicians.

Statistical analysis

Statistical analysis was done using the SPSS packages (SPSS, version 22.0; IBM Corp., Armonk, New York, USA). Normality evaluation was performed by Shapiro-Wilk method. Variables with normally distributed were evaluated by Paired t-test, while for non-normally distributed variables Wilcoxon signed ranks test was applied. Lin concordance correlation coefficient (Lin CCC) was also used for the agreement between the variables. In all measurements, P-value < 0.05 was considered as statistically significant.

RESULTS

[Table 1](#) shows the mean of COV, SNR and LBR in all reconstruction protocols. Although, decreasing the acquisition time led to increased COV for all TOF protocols, no statistically significant differences for COV can be seen between non-TOF and TOF protocols with 18 subsets and 6.4 mm filter size for all scan time (3, 2.5 and 2 min), 18 subsets with 5.4 mm filter size for 3- and 2.5-min scan time and 24 subsets with 6.4 mm filter size for 3 min scan time. Statistically significant improvement in SNR between TOF protocols and non-TOF was observed with 18 subsets, 5.4 mm filter size with 3 min scan time per bed position (P-value=0.00) and 18 subsets, 6.4 mm filter size when 2.5- and 3-min scan time were applied per bed position (P-value≤0.02). The statistical analysis of image quality in terms of LBR showed that all TOF protocols with various scan times produced the significant improvement versus non-TOF protocol.

The findings of SUV agreement in various reconstruction methods are shown in [Table 2](#). The moderate agreement ($0.90 \leq CCC < 0.95$) was illustrated for both SUV_{max} and SUV_{50%} separately, between non-TOF and TOF protocol in all scan time (3, 2.5 and 2 min) with 18 subsets and 5.4 mm post smoothing filter. Although, Substantial agreement ($0.95 \leq CCC < 0.99$) and almost perfect agreement ($CCC \geq 0.99$) was also seen for both SUV_{max} and SUV_{50%} in other protocols.

The relative differences for SUV_{max}, SUV_{50%} and lesion size between all TOF protocols with various scan time and non-TOF are illustrated in [Table 3](#). The smaller variation was seen in TOF protocol for all scan time when 18 subsets and 6.4 mm filter size were applied. Regarding SUV and lesion size variations, it seems that the more impact is seen for smaller filter size (5.4 mm) versus higher subset number (24).

The relative differences for volumetric parameters (TLG_{45%}, MTV_{45%}, TLG_{50%}, MTV_{50%}, TLG_{70%}, MTV_{70%}, TLG_{75%} and MTV_{75%}) between all TOF protocols and non-TOF for various scan time are shown in [Table 4](#). With increasing of threshold value from 45% to 75%, Δ TLG is increased. For 3 min scan time per bed position, in TOF protocol with 18 subsets and 5.4 mm post smoothing filter, Δ TLG varied from 6.77 ± 12.24 to 13.01 ± 14.37 . Although Δ TLG changed also from 9.29 ± 12.76 to 22.45 ± 17.48 in TOF protocol with 24 subsets and 6.4 mm post smoothing filter in 3 min per bed position when threshold varied from 45% to 75%. It is emphasized that the variation of Δ TLG is more emphasized for higher threshold values (70%, 75%) especially when 6.4 mm post smoothing filter was applied.

Although MTV value of TOF protocols is lower than non-TOF protocol in lower threshold value, MTV exhibits higher value versus non-TOF when threshold is increased with more impact for 6.4 mm post smoothing filter. The Δ MTV value were $-10.10 \pm 11.09 (-1.35 \pm 8.59)$, $-7.15 \pm 14.24 (2.30 \pm 12.36)$, $-3.50 \pm 18.57 (10.07 \pm 18.17)$ and $0.68 \pm 17.51 (12.42 \pm 18.90)$ for TOF protocol in 2 min per bed position with 18 subsets and 5.4 mm post smoothing filter (with 18 subsets and 6.4 mm post smoothing filter), when threshold values were 45%, 50%, 70% and 75% respectively.

The coronal and transverse view of CT and PET images of a typical patient with some lesions in lung are shown in [Figure 1](#). Scan time reduction from 3 min to 2 min per bed position cannot significantly cause image degraded even for larger post smoothing filter when higher subset

number (24) with 2 min per bed position was used. It should be clear that there is significant variation for MTV and TLG value separately when threshold was changed from 45% to 75%. It seems that, more variation for both TLG and MTV value is seen for TOF protocols with larger post smoothing filter. TLG_{45%} (MTV_{45%}) were 11525.70(1871.9) and 12158.62(2111.81) for TOF protocol with 18 subsets and 5.4 mm post smoothing filter and TOF protocol with 24 subsets and 6.4 mm post smoothing filter in 2 min per bed position respectively. Although TLG_{75%} (MTV_{75%}) were 3263.22 (391.19) and 3790.16 (488.99) for above reconstruction protocols.

DISCUSSION

Patient motion can produce some error in qualitative and quantitative assessments of PET/CT images [27, 38-40]. Shorter acquisition time can reduce motion artifact especially for patient suffering of pain in oncology PET/CT imaging. It was shown that the evaluation of parameters that related to SUV and volumetric indicators can be important for assessment of high-risk for recurrence after total thyroidectomy in differentiated thyroid cancer patients with primary MTV > 10.0 cm³ [41]. Current work assessed the feasibility of scan time reduction for combination of PSF and TOF protocols (TOF) versus non-TOF protocol (OSEM+PSF), when quantitative including volumetric parameters were evaluated. In general, our study demonstrated that scan time reduction can lead to increase in noise level. In addition, TOF reconstruction with 2 iterations, 18 subsets and 5.4 mm (2min per bed position) and TOF reconstruction with 2 iterations, 24 subsets and 6.4 mm (2.5 and 2 min) introduced significantly high noise level compared to non-TOF. It seems interesting that although SNR improvement can be seen in TOF protocols even in shorter acquisition time, the more caution is needed for TOF protocol with higher subsets (24) especially with shorter scan time (2 min scan time per bed position).

Our findings revealed that smaller filter size and higher subset number produced higher LBR. Our result is in agreement with Sadick et al. [42] and Ferretti et al. [43] that explained higher contrast is seen in application of thinner post-smoothing filter. Previous studies also revealed that combination of PSF and TOF [8, 35, 36, 44] and also TOF alone [8, 35] resulted in higher SUV which is in line with our findings. Scan time reduction from 3 to 2 min per bed position did not exhibit significant variation in SUVs value for individual TOF protocol (data were not shown). However, more variation was seen on SUVs value when small post smoothing filter and higher subset value was applied that is in line with the previous studies[18, 36]. We also stated that there was no agreement between SUV_{max} and SUV_{50%} for all scan time according to Lin concordance correlation coefficient analysis. Our results also showed that all TOF protocols revealed that lesion sizes can be diminished versus non-TOF which is in line with the findings of previous studies [9, 45]; although the lesion size was not changed significantly with decreasing of scan time.

The behavior of MTV can be different when different LBRs and lesion sizes with various threshold values were applied [18, 19]. Previous phantom study [19] revealed that except for SBR=2(with 10mm sphere diameter) and applying of 40 and 50 thresholding value, smaller post smoothing filter can introduce the lower MTV value in all reconstruction protocols that is in agreement with our results. For MTV value, the more variation between TOF and non-TOF in lower threshold (45%, 50%) was observed for 5.4 mm post smoothing filter and in higher threshold (70%, 75%) was seen for 6.4mm post smoothing filter.

Previous studies have suggested that TLG value can be also affected by the SBR, reconstruction algorithm and lesion size, however, this variation is decreased by increasing of lesion size [18] that is in agreement with our findings. It should be clear that the more variation in TLG value between TOF and non-TOF protocol for all threshold value was seen for 6.4 mm post smoothing filter. Our results notify that in all TOF protocols for both TLG and MTV, the values

are decreased with increasing of threshold with significantly more impact for higher threshold (70%, 75%). So as for TOF protocols with 24 subsets, 6.4 mm post smoothing filter size and 3min per bed position, TLG (MTV) were seen 14406.21 ± 12158.98 (3277.97 ± 1902.35), 12359.92 ± 10313.99 (2690.90 ± 1648.53), 5044.37 ± 3774.69 (935.95 ± 625.62) and 3802.36 ± 2627.14 (673.88 ± 419.43) when thresholds were 45%, 50%, 70% and 75% respectively (all data are not shown). Furthermore, higher values for TLG and MTV in all TOF protocols are observed when higher post smoothing filter was applied for each specified threshold value (data are not shown).

It should be emphasized that for each TOF protocols and threshold value, scan time reduction cannot lead to significant variation for both TLG and MTV value. It has been shown that PSF and TOF can improve the image quality [9] and volumetric accuracy [19]. Based on our findings, combination of PSF and TOF protocols (TOF) versus non-TOF protocol (OSEM+PSF) increased TLG values. Our study had some limitations. The lesions did not classify into specified groups according to lesion size and LBR. According to our lesion size (range: 18.31-42.70mm in FWHM) and SBR (range: 1.60-6.76) that evaluated in this study, we concluded that the scan time reduction can be done when combination of TOF and PSF protocol was applied versus non-TOF protocol (PSF only: routine protocol), when administrated activity was not changed (4.6MBq/kg). However, more caution must be taken when higher threshold value was applied especially for higher post smoothing filter and higher subset number.

CONCLUSION

Reduction of scan time utilizing TOF protocol is possible, without affecting measured lesion size and SUV values significantly. However, lower SNRs can be seen on images reconstructed with higher subset number and post smoothing filter for TOF images versus non-TOF ones especially when scan time is decreased. MTV value of TOF protocols is lower than non-TOF protocol in lower threshold value; however, MTV illustrated higher value versus non-TOF when threshold is increased with more impact for 6.4 mm post smoothing filter. TLG for TOF protocols is increased versus non-TOF, when threshold value is increased with more impact for higher threshold value (70%, 75%) especially for 6.4 mm post smoothing filter. In all TOF protocols for both TLG and MTV, the values are decreased with increasing of segmentation threshold, as expected, with significantly more impact for higher threshold (70%, 75%). Furthermore, higher values for TLG and MTV in all TOF protocols were observed when higher post smoothing filter was applied for each specified threshold value. This study showed that scan time reduction from 3 min to 2 min per bed position, for patients that injected with 4.6 MBq/kg of [¹⁸F]FDG with large lesion size (range: 18.31-42.70mm in FWHM) and SBR (range: 1.60-6.76) could be possible in PET/CT images without degrading image quality when combining TOF and PSF protocols relative to non-TOF (PSF only) protocol. However more caution must be exercised for higher subset number (24) and 6.4 mm post smoothing filter size, especially when higher threshold value are utilized.

ACKNOWLEDGMENTS

This work was supported under grant number 36292, Tehran University of Medical Sciences and performed at Masih Daneshvari Hospital in Shahid Beheshti University of Medical Sciences, Tehran, Iran.

REFERENCES

1. Rahmim A, Qi J, Sossi V. Resolution modeling in PET imaging: theory, practice, benefits, and pitfalls. *Med Phys*. 2013 Jun;40(6):064301.

2. Ahangari S, Ghafarian P, Shekari M, Ghadiri H, Bakhshayeshkaram M, Ay MR. The impact of point spread function modeling on scan duration in PET imaging. *Front biomed technol.* 2015 Sep 30;2(3):137-45.
3. Zeraatkar N, Ay MR, Ghafarian P, Sarkar S, Geramifar P, Rahmim A. Monte Carlo-based evaluation of inter-crystal scatter and penetration in the PET subsystem of three GE Discovery PET/CT scanners. *Nucl Instrum Methods Phys Res A.* 2011 Dec 11;659(1):508-14.
4. Tong S, Alessio AM, Kinahan PE. Noise and signal properties in PSF-based fully 3D PET image reconstruction: an experimental evaluation. *Phys Med Biol.* 2010 Mar 7;55(5):1453-73.
5. Hosseini SA, Shiri I, Hajianfar G, Bahadorzadeh B, Ghafarian P, Zaidi H, Ay MR. Synergistic impact of motion and acquisition/reconstruction parameters on ¹⁸F-FDG PET radiomic features in non-small cell lung cancer: Phantom and clinical studies. *Med Phys.* 2022 Jun;49(6):3783-3796.
6. Karp JS, Surti S, Daube-Witherspoon ME, Muehllehner G. Benefit of time-of-flight in PET: experimental and clinical results. *J Nucl Med.* 2008 Mar;49(3):462-70.
7. Kadrmas DJ, Casey ME, Conti M, Jakoby BW, Lois C, Townsend DW. Impact of time-of-flight on PET tumor detection. *J Nucl Med.* 2009 Aug;50(8):1315-23.
8. Shekari M, Ghafarian P, Ahangari S, Ay MR. Quantification of the impact of TOF and PSF on PET images using the noise-matching concept: clinical and phantom study. *Nucl Sci Tech.* 2017 Nov;28(11):1-8.
9. Sharifpour R, Ghafarian P, Rahmim A, Ay MR. Quantification and reduction of respiratory induced artifacts in positron emission tomography/computed tomography using the time-of-flight technique. *Nucl Med Commun.* 2017 Nov;38(11):948-955.
10. Hemmati H, Kamali-Asl A, Ay M, Ghafarian P. Compton scatter tomography in TOF-PET. *Phys Med Biol.* 2017 Sep 12;62(19):7641-7658.
11. Surti S, Karp JS. Experimental evaluation of a simple lesion detection task with time-of-flight PET. *Phys Med Biol.* 2009 Jan 21;54(2):373-84. doi: 10.1088/0031-9155/54/2/013. Epub 2008 Dec 19. Erratum in: *Phys Med Biol.* 2009 Feb 21;54(4):1087.
12. Schaefferkoetter J, Casey M, Townsend D, El Fakhri G. Clinical impact of time-of-flight and point response modeling in PET reconstructions: a lesion detection study. *Phys Med Biol.* 2013 Mar 7;58(5):1465-78.
13. Conti M. Focus on time-of-flight PET: the benefits of improved time resolution. *Eur J Nucl Med Mol Imaging.* 2011 Jun;38(6):1147-57.
14. Shekari M, Ghafarian P, Ahangari S, Ghadiri H, Bakhshayeshkaram M, Ay MR. Optimizing image reconstruction parameters in time of flight PET/CT imaging: a phantom study. *Front biomed technol.* 2015 Sep 30;2(3):146-54.
15. Moon SH, Choi JY, Lee HJ, Son YI, Baek CH, Ahn YC, Park K, Lee KH, Kim BT. Prognostic value of ¹⁸F-FDG PET/CT in patients with squamous cell carcinoma of the tonsil: comparisons of volume-based metabolic parameters. *Head Neck.* 2013 Jan;35(1):15-22.
16. Boellaard R. Standards for PET image acquisition and quantitative data analysis. *J Nucl Med.* 2009 May;50 Suppl 1:11S-20S.
17. Nahmias C, Wahl LM. Reproducibility of standardized uptake value measurements determined by ¹⁸F-FDG PET in malignant tumors. *J Nucl Med.* 2008 Nov;49(11):1804-8.
18. Ketabi A, Ghafarian P, Mosleh-Shirazi MA, Mahdavi SR, Ay MR. The influence of using different reconstruction algorithms on sensitivity of quantitative ¹⁸F-FDG-PET volumetric measures to background activity variation. *Iran J Nucl Med.* 2018;26(2):87-97.

19. Ketabi A, Ghafarian P, Mosleh-Shirazi MA, Mahdavi SR, Rahmim A, Ay MR. Impact of image reconstruction methods on quantitative accuracy and variability of FDG-PET volumetric and textural measures in solid tumors. *Eur Radiol.* 2019 Apr;29(4):2146-2156.
20. Paidpally V, Chirindel A, Chung CH, Richmon J, Koch W, Quon H, Subramaniam RM. FDG volumetric parameters and survival outcomes after definitive chemoradiotherapy in patients with recurrent head and neck squamous cell carcinoma. *AJR Am J Roentgenol.* 2014 Aug;203(2):W139-45.
21. Ryu IS, Kim JS, Roh JL, Cho KJ, Choi SH, Nam SY, Kim SY. Prognostic significance of preoperative metabolic tumour volume and total lesion glycolysis measured by (18)F-FDG PET/CT in squamous cell carcinoma of the oral cavity. *Eur J Nucl Med Mol Imaging.* 2014 Mar;41(3):452-61.
22. Teimourian B, Ay MR, Zafarghandi MS, Ghafarian P, Ghadiri H, Zaidi H. A novel energy mapping approach for CT-based attenuation correction in PET. *Med Phys.* 2012 Apr;39(4):2078-89.
23. Sridhar P, Mercier G, Tan J, Truong MT, Daly B, Subramaniam RM. FDG PET metabolic tumor volume segmentation and pathologic volume of primary human solid tumors. *AJR Am J Roentgenol.* 2014 May;202(5):1114-9.
24. Zaidi H, El Naqa I. PET-guided delineation of radiation therapy treatment volumes: a survey of image segmentation techniques. *Eur J Nucl Med Mol Imaging.* 2010 Nov;37(11):2165-87.
25. Rogasch JM, Hofheinz F, Lougovski A, Furth C, Ruf J, Großer OS, Mohnike K, Hass P, Walke M, Amthauer H, Steffen IG. The influence of different signal-to-background ratios on spatial resolution and F18-FDG-PET quantification using point spread function and time-of-flight reconstruction. *EJNMMI Phys.* 2014 Dec;1(1):12.
26. Knudtsen IS, van Elmpt W, Ollers M, Malinen E. Impact of PET reconstruction algorithm and threshold on dose painting of non-small cell lung cancer. *Radiother Oncol.* 2014 Nov;113(2):210-4.
27. Rezaei S, Ghafarian P, Jha AK, Rahmim A, Sarkar S, Ay MR. Joint compensation of motion and partial volume effects by iterative deconvolution incorporating wavelet-based denoising in oncologic PET/CT imaging. *Phys Med.* 2019 Dec;68:52-60.
28. Rezaei S, Ghafarian P, Bakhshayesh-Karam M, Uribe CF, Rahmim A, Sarkar S, Ay MR. The impact of iterative reconstruction protocol, signal-to-background ratio and background activity on measurement of PET spatial resolution. *Jpn J Radiol.* 2020 Mar;38(3):231-239.
29. Albano D, Re A, Tucci A, Giubbini R, Bertagna F. Prognostic role of Δ MTV and Δ TLG in Burkitt lymphoma. *Ann Nucl Med.* 2019 Apr;33(4):280-287.
30. Øen SK, Aasheim LB, Eikenes L, Karlberg AM. Image quality and detectability in Siemens Biograph PET/MRI and PET/CT systems-a phantom study. *EJNMMI Phys.* 2019 Aug 5;6(1):16.
31. Sharma A, Mohan A, Bhalla AS, Sharma MC, Vishnubhatla S, Das CJ, Pandey AK, Sekhar Bal C, Patel CD, Sharma P, Agarwal KK, Kumar R. Role of various metabolic parameters derived from baseline 18F-FDG PET/CT as prognostic markers in non-small cell lung cancer patients undergoing platinum-based chemotherapy. *Clin Nucl Med.* 2018 Jan;43(1):e8-e17.
32. Gencturk M, Ozturk K, Koxsel Y, Li F, Cayci Z. Pretreatment quantitative ¹⁸F-FDG PET/CT parameters as a predictor of survival in adenoid cystic carcinoma of the salivary glands. *Clin Imaging.* 2019 Jan-Feb;53:17-24.
33. Ye S, Liu S, Zhou S, Xiang L, Wu X, Yang H. The role of 18F-FDG PET/CT-based quantitative metabolic parameters in patients with ovarian clear cell carcinoma. *Cancer Biomark.* 2020;27(2):189-194.

34. Prieto E, Domínguez-Prado I, García-Velloso MJ, Peñuelas I, Richter JÁ, Martí-Climent JM. Impact of time-of-flight and point-spread-function in SUV quantification for oncological PET. *Clin Nucl Med*. 2013 Feb;38(2):103-9.
35. Akamatsu G, Mitsumoto K, Taniguchi T, Tsutsui Y, Baba S, Sasaki M. Influences of point-spread function and time-of-flight reconstructions on standardized uptake value of lymph node metastases in FDG-PET. *Eur J Radiol*. 2014 Jan;83(1):226-30.
36. Sharifpour R, Ghafarian P, Bakhshayesh-Karam M, Jamaati H, Ay MR. Impact of time-of-flight and point-spread-function for respiratory artifact reduction in PET/CT imaging: focus on standardized uptake value. *Tanaffos*. 2017;16(2):127-135.
37. Boellaard R, Delgado-Bolton R, Oyen WJ, Giammarile F, Tatsch K, Eschner W, Verzijlbergen FJ, Barrington SF, Pike LC, Weber WA, Stroobants S, Delbeke D, Donohoe KJ, Holbrook S, Graham MM, Testanera G, Hoekstra OS, Zijlstra J, Visser E, Hoekstra CJ, Pruim J, Willemsen A, Arends B, Kotzerke J, Bockisch A, Beyer T, Chiti A, Krause BJ. FDG PET/CT: EANM procedure guidelines for tumour imaging: version 2.0. *Eur J Nucl Med Mol Imaging*. 2015 Feb;42(2):328-54.
38. Ghafarian P, Ay MR, Fard-Esfahani A, Rahmim A, Zaidi H. Quantification of PET and CT misalignment errors due to bulk motion in cardiac PET/CT imaging: phantom and clinical and studies. *Front biomed technol*. 2014;1(3):159-67.
39. Ghafarian P, Ay MR. The influence of PET and CT misalignment due to respiratory motion on the cardiac PET/CT imaging: a simulation study. *Front biomed technol*. 2014 Dec 30;1(4):252-7.
40. Bettinardi V, Picchio M, Di Muzio N, Gilardi MC. Motion management in positron emission tomography/computed tomography for radiation treatment planning. *Semin Nucl Med*. 2012 Sep;42(5):289-307.
41. Nakajo M, Jinguji M, Shinaji T, Tani A, Nakabeppu Y, Nakajo M, Nakajo A, Natsugoe S, Yoshiura T. ¹⁸F-FDG-PET/CT features of primary tumours for predicting the risk of recurrence in thyroid cancer after total thyroidectomy: potential usefulness of combination of the SUV-related, volumetric, and heterogeneous texture parameters. *Br J Radiol*. 2019 Feb;92(1094):20180620.
42. Sadick M, Molina F, Frey S, Piniol R, Sadick H, Brade J, Fink C, Schoenberg SO, He Y. Effect of reconstruction parameters in high-definition PET/CT on assessment of lymph node metastases in head and neck squamous cell carcinoma. *J Nucl Med Technol*. 2013 Mar;41(1):19-25.
43. Ferretti A, Bellan E, Gava M, Chondrogiannis S, Massaro A, Nibale O, Rubello D. Phantom study of the impact of reconstruction parameters on the detection of mini- and micro-volume lesions with a low-dose PET/CT acquisition protocol. *Eur J Radiol*. 2012 Nov;81(11):3363-70.
44. Sheikhabaei S, Marcus C, Wray R, Rahmim A, Lodge MA, Subramaniam RM. Impact of point spread function reconstruction on quantitative ¹⁸F-FDG-PET/CT imaging parameters and inter-reader reproducibility in solid tumors. *Nucl Med Commun*. 2016 Mar;37(3):288-96.
45. Delso G, Khalighi M, Ter Voert E, Barbosa F, Sekine T, Hüllner M, Veit-Haibach P. Effect of time-of-flight information on PET/MR reconstruction artifacts: comparison of free-breathing versus breath-hold MR-based attenuation correction. *Radiology*. 2017 Jan;282(1):229-235.

Figure Legend

Fig 1. comparison of coronal (a) and transverse (b) views of reconstructed CT and PET images of a typical patient (70 Kg weight and administration of 328 MBq of [¹⁸F]FDG with some abnormal lesions in lung region for different scan time per bed position. From left to right: CT, non-TOF, TOF^a_3min , TOF^a_2min, TOF^b_3min , TOF^b_2min, TOF^c_3min and TOF^c_2min; non-TOF(routine protocol; 3it, 18subset,6.4mm filter), TOF^a(2it, 18subset,5.4mm filter), TOF^b(2it, 18subset,6.4mm filter)and TOF^c(2it, 24subset, 6.4mm filter).

Table 1: The mean±SD value of image quality parameters and P-value between TOF and non-TOF protocols.

Reconstruction protocol	COV		SNR		LBR	
	mean±SD	P-Value	mean±SD	P-Value	mean±SD	P-Value
non-TOF ^a	6.79±0.55		28.12±18.91		2.97±1.40	
TOF ^b _3 min	6.79±0.66	0.28	34.92±18.49	0.00	3.53±1.38	0.00
TOF ^b _2.5 min	7.31±0.90	0.08	31.93±16.16	0.09	3.53±1.36	0.00
TOF ^b _2 min	8.08±1.25	0.02	28.42±14.31	0.90	3.53±1.39	0.00
TOF ^c _3 min	6.05±0.58	0.50	35.61±19.54	0.00	3.31±1.62	0.01
TOF ^c _2.5 min	6.49±0.79	0.65	32.75±17.26	0.02	3.31±1.58	0.01
TOF ^c _2 min	7.19±1.12	0.12	28.96±15.18	0.67	3.30±1.57	0.01
TOF ^d _3 min	7.42±0.81	0.05	30.18±15.86	0.24	3.42±1.43	0.00
TOF ^d _2.5 min	7.89±1.00	0.03	28.30±14.21	0.94	3.42±1.40	0.00
TOF ^d _2 min	8.75±1.14	0.01	25.56±12.72	0.36	3.43±1.40	0.00

- a. 3it, 18 subset, 6.4 mm filter.
- b. 2it, 18 subset, 5.4 mm filter.
- c. 2it, 18 subset, 6.4 mm filter.
- d. 2it, 24 subset, 6.4 mm filter.

Table 2: Lin Concordance Correlation Coefficients between SUV measured with different protocols.

		SUVmax										SUV50%									
		non-TOF ^a	TOF ^b _3 min	TOF ^b _2.5 min	TOF ^b _2 min	TOF ^c _3 min	TOF ^c _2.5 min	TOF ^c _2 min	TOF ^d _3 min	TOF ^d _2.5 min	TOF ^d _2 min	non-TOF ^a	TOF ^b _3 min	TOF ^b _2.5 min	TOF ^b _2 min	TOF ^c _3 min	TOF ^c _2.5 min	TOF ^c _2 min	TOF ^d _3 min	TOF ^d _2.5 min	TOF ^d _2 min
SUVmax	non-TOF ^a	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---
	TOF ^b _3min	0.94	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---
	TOF ^b _2.5 min	0.94	1.00	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---
	TOF ^b _2min	0.94	1.00	1.00	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---
	TOF ^c _3min	0.98	0.99	0.99	0.99	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---
	TOF ^c _2.5 min	0.98	0.98	0.99	0.99	1.00	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---
	TOF ^c _2min	0.98	0.98	0.98	0.99	1.00	1.00	---	---	---	---	---	---	---	---	---	---	---	---	---	---
	TOF ^d _3min	0.97	0.99	0.99	0.99	1.00	1.00	1.00	---	---	---	---	---	---	---	---	---	---	---	---	---
	TOF ^d _2.5 min	0.97	0.99	0.99	0.99	1.00	1.00	1.00	1.00	---	---	---	---	---	---	---	---	---	---	---	---
	TOF ^d _2min	0.97	0.98	0.99	0.99	0.99	0.99	1.00	1.00	1.00	---	---	---	---	---	---	---	---	---	---	---
SUV50%	non-TOF ^a	0.70	0.78	0.55	0.55	0.62	0.62	0.62	0.59	0.59	0.59	---	---	---	---	---	---	---	---	---	---
	TOF ^b _3min	0.84	0.70	0.71	0.71	0.77	0.78	0.78	0.74	0.74	0.74	0.93	---	---	---	---	---	---	---	---	---
	TOF ^b _2.5 min	0.83	0.70	0.70	0.70	0.77	0.78	0.77	0.74	0.74	0.74	0.93	1.00	---	---	---	---	---	---	---	---
	TOF ^b _2min	0.83	0.69	0.69	0.70	0.76	0.77	0.77	0.73	0.73	0.73	0.93	1.00	1.00	---	---	---	---	---	---	---
	TOF ^c _3min	0.77	0.63	0.63	0.63	0.70	0.70	0.70	0.67	0.67	0.67	0.97	0.99	0.99	0.99	---	---	---	---	---	---
	TOF ^c _2.5 min	0.76	0.62	0.63	0.63	0.69	0.70	0.70	0.66	0.66	0.66	0.97	0.98	0.99	0.99	1.00	---	---	---	---	---
	TOF ^c _2min	0.76	0.61	0.62	0.62	0.69	0.69	0.69	0.66	0.66	0.66	0.97	0.98	0.98	0.99	1.00	1.00	---	---	---	---
	TOF ^d _3min	0.79	0.64	0.64	0.65	0.71	0.72	0.72	0.68	0.69	0.68	0.96	0.99	0.99	0.99	1.00	1.00	1.00	---	---	---
	TOF ^d _2.5 min	0.78	0.64	0.64	0.64	0.71	0.71	0.71	0.68	0.68	0.68	0.96	0.99	0.99	0.99	1.00	1.00	1.00	1.00	---	---
	TOF ^d _2min	0.78	0.63	0.64	0.64	0.71	0.71	0.71	0.68	0.68	0.68	0.96	0.99	0.98	0.99	0.99	0.99	1.00	1.00	1.00	---

Lin Concordance Correlation Coefficients (CCC) greater than 0.95 are in bold.

- a. 3it, 18 subset, 6.4 mm filter.
- b. 2it, 18 subset, 5.4 mm filter.
- c. 2it, 18 subset, 6.4 mm filter.
- d. 2it, 24 subset, 6.4 mm filter.

Table 3: The relative difference for two types of SUVs and lesion size between TOF and non-TOF protocols .

Reconstruction protocol	ΔSUVmax	ΔSUV50%	ΔFWHM
TOF ^a _3 min	20.11±9.87	20.59±10.57	-10.95±7.08
TOF ^a _2.5 min	20.63±10.07	21.77±10.73	-10.46±7.88
TOF ^a _2 min	20.90±11.39	21.20±11.50	-10.71±9.85
TOF ^b _3 min	12.35±7.86	12.81±8.23	-5.70±7.90
TOF ^b _2.5 min	12.82±7.79	13.61±8.35	-5.55±8.43
TOF ^b _2 min	12.92±9.39	13.70±9.43	-5.27±9.98
TOF ^c _3 min	16.52±9.39	16.71±10.50	-10.17±7.91
TOF ^c _2.5 min	17.30±11.21	18.09±11.12	-9.99±8.20
TOF ^c _2 min	18.73±15.53	19.51±12.74	-10.54±10.85

a. 2it, 18 subset, 5.4 mm filter.


b. 2it, 18 subset, 6.4 mm filter.

c. 2it, 24 subset, 6.4 mm filter.


Table 4: The relative difference in four types of TLGs between TOF and non-TOF protocols.

Reconstruction protocol	Δ TLG45%	Δ MTV45%	Δ TLG50%	Δ MTV50%	Δ TLG70%	Δ MTV70%	Δ TLG75%	Δ MTV75%
TOF ^a _3 min	6.77±12.24	-10.39±12.14	7.56±13.17	-9.02±16.20	11.99±15.51	-4.97±17.15	13.01±14.37	-2.68±15.62
TOF ^b _2.5 min	7.38±11.22	-9.90±12.18	7.65±11.32	-9.63±15.08	14.38±12.10	-3.59±14.68	16.79±15.01	1.25±18.56
TOF ^a _2 min	6.35±9.49	-10.10±11.09	8.51±10.0	-7.15±14.24	13.80±15.93	-3.50±18.57	17.37±16.20	0.68±17.51
TOF ^b _3 min	11.70±12.74	-1.26±11.36	12.70±13.61	0.92±13.90	16.17±17.52	5.08±20.12	22.79±13.98	10.81±14.72
TOF ^b _2.5 min	12.47±11.37	-1.22±10.63	13.42±11.92	0.60±13.53	20.19±14.92	6.84±19.78	22.75±13.18	11.12±15.97
TOF ^b _2 min	12.15±11.01	-1.35±8.59	12.90±10.28	2.30±12.36	21.09±13.45	10.07±18.17	23.30±16.40	12.42±18.90
TOF ^c _3 min	9.29±12.76	-5.60±11.03	11.31±13.43	-1.98±14.45	13.98±15.10	-1.13±18.83	22.45±17.48	8.44±18.00
TOF ^c _2.5 min	10.32±11.50	-4.33±11.17	11.53±10.46	-2.88±12.55	18.02±14.65	1.08±18.89	26.80±17.89	12.34±18.96
TOF ^c _2 min	9.50±9.87	-4.81±11.29	10.11±9.63	-3.97±12.19	16.54±16.74	0.35±20.95	21.28±20.17	7.83±23.93

a. 2it, 18subset, 5.4mm filter.
b. 2it, 18subset, 6.4mm filter.
c. 2it, 24subset, 6.4mm filter.



(a)



(b)